

EC Type-Examination Certificate

Number 3694 Issue 2

Date of first Issue: 30th May 2008

This is to certify that the product group reference "Household Gloves" comprising the following products:

Product Reference	Description	Classification:			
Lightweight Household Gloves	Five fingered flock lined latex gloves.	EN 388:2003	Level	EN 374-3:2003	Permeation level
Colours: Green, Red, Yellow, Pink, Orange, Blue.		Abrasion resistance	0	n-Heptane	0
Sizes 7-10		Blade cut resistance	0	Sulphuric Acid 96%	1
		Tear resistance	1	Sodium Hydroxide 40%	5
		Puncture resistance	0		
				EN 374-2:2003	
				Air leak	Pass
				Water Leak	Pass

Technical Reports:

SATRA: SPC0159457/0801/DMcK/DC, SPC0159457/0801/IG,
 CHM0228718/1440/EN/A
 FC-BIOS : 108432/TEST SOP 110/080421


has been subject to an EC Type-examination in accordance with Article 10 of the PPE Directive (89/686/EEC) and has been shown to satisfy the relevant provisions of this Directive for the complex category through:

- i Testing to the following standard: EN 374-1:2003 (Excluding Clause 5.3.2), EN388:2003
- ii Examination of the relevant technical documentation.

You are therefore licensed to mark the product(s) listed above in accordance with Article 13 of Directive (89/686/EEC) and any relevant amending Directives once you have drawn up an EC declaration of product conformity. Please note that:

1. Full details of the certification and product are contained in the manufacturer's technical file
2. This certificate is only valid if embossed with the text – "SATRA – European Notified Body 0321"
3. This certificate is issued subject to the conditions on the reverse side of this certificate
4. CE marking of production is also reliant on current compliance with Directive 89/686/EEC Article 11

Signed:  (T. Holly)

Signed: 
 On behalf of SATRA

Date: 31st October 2014